



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092222\
Data File : VX031588.D
Acq On : 22 Sep 2022 22:41
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
Title : SW846 8260
Last Update : Tue Sep 20 09:45:29 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	23.7	3860	PASS
75	95	30	60	59.4	9675	PASS
95	95	100	100	100.0	16278	PASS
96	95	5	9	7.2	1179	PASS
173	174	0.00	2	1.6	235	PASS
174	95	50	100	91.3	14860	PASS
175	174	5	9	7.6	1134	PASS
176	174	95	101	96.6	14355	PASS
177	176	5	9	6.8	976	PASS