

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092223\
 Data File : VX037864.D
 Acq On : 22 Sep 2023 09:17
 Operator : JC/MD
 Sample : VX0922MBL01
 Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0922MBL01

Quant Time: Sep 23 01:19:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	227565	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	379407	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	362968	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	185417	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	137635	46.034	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.060%
35) Dibromofluoromethane	5.385	113	118809	46.800	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.600%
50) Toluene-d8	8.647	98	459783	48.744	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.480%
62) 4-Bromofluorobenzene	11.079	95	168901	49.276	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.560%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092223\
Data File : VX037864.D
Acq On : 22 Sep 2023 09:17
Operator : JC/MD
Sample : VX0922MBL01
Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0922MBL01

Quant Time: Sep 23 01:19:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 16 04:55:44 2023
Response via : Initial Calibration

