

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092223\
 Data File : VX037863.D
 Acq On : 22 Sep 2023 08:43
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 23 01:18:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	97	0.00
2 T	Dichlorodifluoromethane	0.527	0.470	10.8	88	0.00
3 P	Chloromethane	0.573	0.491	14.3	90	0.00
4 C	Vinyl Chloride	0.619	0.549	11.3#	88	0.00
5 T	Bromomethane	0.443	0.441	0.5	96	0.00
6 T	Chloroethane	0.426	0.381	10.6	90	0.00
7 T	Trichlorofluoromethane	0.995	0.947	4.8	92	0.00
8 T	Diethyl Ether	0.385	0.350	9.1	91	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.505	0.494	2.2	96	0.00
10 T	Methyl Iodide	0.668	0.583	12.7	82	0.00
11 T	Tert butyl alcohol	0.141	0.108	23.4	77	-0.01
12 CM	1,1-Dichloroethene	0.488	0.455	6.8#	90	0.00
13 T	Acrolein	0.139	0.120	13.7	86	0.00
14 T	Allyl chloride	0.812	0.738	9.1	89	0.00
15 T	Acrylonitrile	0.314	0.280	10.8	87	0.00
16 T	Acetone	0.316	0.278	12.0	93	0.00
17 T	Carbon Disulfide	1.223	1.069	12.6	85	0.00
18 T	Methyl Acetate	1.140	1.030	9.6	90	0.00
19 T	Methyl tert-butyl Ether	1.794	1.679	6.4	93	-0.01
20 T	Methylene Chloride	0.641	0.555	13.4	93	0.00
21 T	trans-1,2-Dichloroethene	0.542	0.511	5.7	91	0.00
22 T	Diisopropyl ether	1.692	1.609	4.9	93	0.00
23 T	Vinyl Acetate	1.245	1.181	5.1	90	0.00
24 P	1,1-Dichloroethane	1.014	0.947	6.6	92	0.00
25 T	2-Butanone	0.446	0.393	11.9	87	-0.02
26 T	2,2-Dichloropropane	0.819	0.803	2.0	97	0.00
27 T	cis-1,2-Dichloroethene	0.664	0.629	5.3	92	0.00
28 T	Bromochloromethane	0.473	0.419	11.4	90	0.00
29 T	Tetrahydrofuran	0.277	0.237	14.4	83	-0.02
30 C	Chloroform	1.092	1.040	4.8#	93	0.00
31 T	Cyclohexane	0.843	0.791	6.2	89	0.00
32 T	1,1,1-Trichloroethane	0.925	0.902	2.5	92	0.00
33 S	1,2-Dichloroethane-d4	0.657	0.654	0.5	97	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	94	0.00
35 S	Dibromofluoromethane	0.335	0.355	-6.0	99	0.00
36 T	1,1-Dichloropropene	0.461	0.455	1.3	93	0.00
37 T	Ethyl Acetate	0.507	0.453	10.7	84	-0.02
38 T	Carbon Tetrachloride	0.482	0.492	-2.1	94	0.00
39 T	Methylcyclohexane	0.585	0.587	-0.3	94	0.00
40 TM	Benzene	1.417	1.366	3.6	91	0.00
41 T	Methacrylonitrile	0.258	0.242	6.2	88	-0.03
42 TM	1,2-Dichloroethane	0.512	0.494	3.5	93	-0.01
43 T	Isopropyl Acetate	0.810	0.760	6.2	87	0.00
44 TM	Trichloroethene	0.395	0.372	5.8	93	-0.01
45 C	1,2-Dichloropropane	0.368	0.357	3.0#	93	0.00
46 T	Dibromomethane	0.265	0.261	1.5	92	0.00
47 T	Bromodichloromethane	0.488	0.506	-3.7	95	0.00
48 T	Methyl methacrylate	0.387	0.370	4.4	88	-0.01

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092223\
 Data File : VX037863.D
 Acq On : 22 Sep 2023 08:43
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 23 01:18:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.010	0.008	20.0	78	-0.02
50 S	Toluene-d8	1.243	1.319	-6.1	97	0.00
51 T	4-Methyl-2-Pentanone	0.514	0.474	7.8	86	0.00
52 CM	Toluene	0.913	0.899	1.5#	91	0.00
53 T	t-1,3-Dichloropropene	0.510	0.545	-6.9	95	-0.01
54 T	cis-1,3-Dichloropropene	0.564	0.586	-3.9	94	0.00
55 T	1,1,2-Trichloroethane	0.385	0.381	1.0	92	0.00
56 T	Ethyl methacrylate	0.575	0.568	1.2	90	0.00
57 T	1,3-Dichloropropane	0.623	0.611	1.9	93	0.00
58 T	2-Chloroethyl Vinyl ether	0.268	0.251	6.3	89	0.00
59 T	2-Hexanone	0.400	0.385	3.8	90	0.00
60 T	Dibromochloromethane	0.384	0.408	-6.2	94	0.00
61 T	1,2-Dibromoethane	0.408	0.407	0.2	91	0.00
62 S	4-Bromofluorobenzene	0.452	0.515	-13.9	106	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	95	0.00
64 T	Tetrachloroethene	0.341	0.327	4.1	94	0.00
65 PM	Chlorobenzene	1.079	1.027	4.8	94	0.00
66 T	1,1,1,2-Tetrachloroethane	0.382	0.386	-1.0	93	0.00
67 C	Ethyl Benzene	1.870	1.855	0.8#	95	0.00
68 T	m/p-Xylenes	0.723	0.722	0.1	94	0.00
69 T	o-Xylene	0.725	0.705	2.8	92	0.00
70 T	Styrene	1.163	1.211	-4.1	96	0.00
71 P	Bromoform	0.302	0.316	-4.6	95	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	111	0.00
73 T	Isopropylbenzene	3.668	3.265	11.0	96	0.00
74 T	N-amyl acetate	1.486	1.297	12.7	92	0.00
75 P	1,1,2,2-Tetrachloroethane	1.290	1.121	13.1	97	0.00
76 T	1,2,3-Trichloropropane	1.156	1.008	12.8	95	0.00
77 T	Bromobenzene	0.908	0.819	9.8	99	0.00
78 T	n-propylbenzene	4.143	3.933	5.1	100	0.00
79 T	2-Chlorotoluene	2.554	2.283	10.6	98	0.00
80 T	1,3,5-Trimethylbenzene	3.098	2.851	8.0	98	0.00
81 T	trans-1,4-Dichloro-2-butene	0.361	0.332	8.0	95	0.00
82 T	4-Chlorotoluene	2.876	2.691	6.4	101	0.00
83 T	tert-Butylbenzene	3.131	2.836	9.4	96	0.00
84 T	1,2,4-Trimethylbenzene	3.082	2.857	7.3	99	0.00
85 T	sec-Butylbenzene	3.798	3.694	2.7	102	0.00
86 T	p-Isopropyltoluene	3.165	3.096	2.2	103	0.00
87 T	1,3-Dichlorobenzene	1.692	1.604	5.2	105	0.00
88 T	1,4-Dichlorobenzene	1.732	1.632	5.8	108	0.00
89 T	n-Butylbenzene	2.772	2.957	-6.7	111	0.00
90 T	Hexachloroethane	0.536	0.531	0.9	103	0.00
91 T	1,2-Dichlorobenzene	1.728	1.610	6.8	106	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.257	0.224	12.8	92	0.00
93 T	1,2,4-Trichlorobenzene	1.018	1.068	-4.9	114	0.00
94 T	Hexachlorobutadiene	0.475	0.497	-4.6	114	0.00
95 T	Naphthalene	3.442	3.260	5.3	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092223\
Data File : VX037863.D
Acq On : 22 Sep 2023 08:43
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Sep 23 01:18:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 16 04:55:44 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.047	1.049	-0.2	110	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6