



BFB

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092223\
Data File : VX037862.D
Acq On : 22 Sep 2023 08:13
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Title : SW846 8260
Last Update : Sat Sep 16 04:55:44 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.8	13662	PASS
75	95	30	60	51.9	39800	PASS
95	95	100	100	100.0	76699	PASS
96	95	5	9	6.8	5200	PASS
173	174	0.00	2	1.1	651	PASS
174	95	50	100	79.7	61131	PASS
175	174	5	9	7.3	4455	PASS
176	174	95	101	95.8	58573	PASS
177	176	5	9	6.5	3825	PASS