

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092225\
 Data File : VX047693.D
 Acq On : 22 Sep 2025 11:21
 Operator : JC/MD
 Sample : Q3097-08DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE125D2-20250911DL

Quant Time: Sep 23 01:30:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	210054	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	374235	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	336300	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	163574	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	174560	52.208	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 104.420%	
35) Dibromofluoromethane	5.367	113	131212	52.279	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.560%	
50) Toluene-d8	8.629	98	457155	52.640	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 105.280%	
62) 4-Bromofluorobenzene	11.061	95	174373	52.906	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 105.820%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	3008	1.238	ug/l	95
44) Trichloroethene	7.110	130	144794	53.767	ug/l	95
64) Tetrachloroethene	9.256	164	4669	2.130	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092225\
Data File : VX047693.D
Acq On : 22 Sep 2025 11:21
Operator : JC/MD
Sample : Q3097-08DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE125D2-20250911DL

Quant Time: Sep 23 01:30:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

