

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092225\
 Data File : VX047696.D
 Acq On : 22 Sep 2025 12:23
 Operator : JC/MD
 Sample : Q3097-14DL 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE105D2-20250911DL

Quant Time: Sep 23 01:32:17 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	199018	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	365741	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	330943	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	159199	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	170350	53.774	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.540%	
35) Dibromofluoromethane	5.379	113	128743	52.487	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.980%	
50) Toluene-d8	8.634	98	450242	53.048	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 106.100%	
62) 4-Bromofluorobenzene	11.061	95	171071	53.109	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 106.220%	
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.337	101	1323	0.575	ug/l	95
44) Trichloroethene	7.116	130	156901	59.616	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092225\
Data File : VX047696.D
Acq On : 22 Sep 2025 12:23
Operator : JC/MD
Sample : Q3097-14DL 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE105D2-20250911DL

Quant Time: Sep 23 01:32:17 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

