

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX092225\  
 Data File : VX047697.D  
 Acq On : 22 Sep 2025 12:44  
 Operator : JC/MD  
 Sample : Q3097-09  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 RE108D1-20250911

Quant Time: Sep 23 01:32:42 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X091625W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 17 06:39:58 2025  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units      | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|------------|----------|
| -----                       |        |       |          |          |            |          |
| Internal Standards          |        |       |          |          |            |          |
| 1) Pentafluorobenzene       | 5.538  | 168   | 188813   | 50.000   | ug/l       | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.751  | 114   | 347701   | 50.000   | ug/l       | 0.00     |
| 63) Chlorobenzene-d5        | 10.037 | 117   | 313912   | 50.000   | ug/l       | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.006 | 152   | 152363   | 50.000   | ug/l       | 0.00     |
| System Monitoring Compounds |        |       |          |          |            |          |
| 33) 1,2-Dichloroethane-d4   | 5.940  | 65    | 156802   | 52.173   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 78 - 117 | Recovery | = 104.340% |          |
| 35) Dibromofluoromethane    | 5.373  | 113   | 118807   | 50.949   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | = 101.900% |          |
| 50) Toluene-d8              | 8.635  | 98    | 415067   | 51.441   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 92 - 112 | Recovery | = 102.880% |          |
| 62) 4-Bromofluorobenzene    | 11.067 | 95    | 157359   | 51.387   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | = 102.780% |          |
| Target Compounds            |        |       |          |          |            |          |
|                             |        |       |          |          | Qvalue     |          |
| 44) Trichloroethene         | 7.110  | 130   | 60076    | 24.011   | ug/l       | 95       |
| 64) Tetrachloroethene       | 9.262  | 164   | 3707     | 1.811    | ug/l       | 91       |
| -----                       |        |       |          |          |            |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX092225\  
Data File : VX047697.D  
Acq On : 22 Sep 2025 12:44  
Operator : JC/MD  
Sample : Q3097-09  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 10 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
RE108D1-20250911

Quant Time: Sep 23 01:32:42 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X091625W.M  
Quant Title : SW846 8260  
QLast Update : Wed Sep 17 06:39:58 2025  
Response via : Initial Calibration

