

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092225\
 Data File : VX047700.D
 Acq On : 22 Sep 2025 13:46
 Operator : JC/MD
 Sample : PB169755TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB169755TB

Quant Time: Sep 23 01:33:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	146155	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	299797	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	289183	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	140425	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	113788	48.911	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.820%
35) Dibromofluoromethane	5.379	113	93805	46.655	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.320%
50) Toluene-d8	8.635	98	329401	47.348	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.700%
62) 4-Bromofluorobenzene	11.061	95	136757	51.795	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.600%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.947	59	1495	5.770	ug/l #	84
16) Acetone	2.374	43	61775	61.071	ug/l	95
18) Methyl Acetate	2.703	43	14294	6.350	ug/l	99
20) Methylene Chloride	2.794	84	15436	7.209	ug/l	95
25) 2-Butanone	4.556	43	14590	11.563	ug/l	100
37) Ethyl Acetate	4.715	43	9754	3.101	ug/l #	93
43) Isopropyl Acetate	6.336	43	11158	2.163	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092225\
Data File : VX047700.D
Acq On : 22 Sep 2025 13:46
Operator : JC/MD
Sample : PB169755TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB169755TB

Quant Time: Sep 23 01:33:58 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

