

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092225\
 Data File : VX047701.D
 Acq On : 22 Sep 2025 14:07
 Operator : JC/MD
 Sample : Q3133-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VNJ-238

Quant Time: Sep 23 01:34:30 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	158169	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	331918	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	321971	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	163047	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	125373	49.798	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.600%
35) Dibromofluoromethane	5.373	113	103473	46.483	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.960%
50) Toluene-d8	8.635	98	362643	47.081	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.160%
62) 4-Bromofluorobenzene	11.061	95	153483	52.505	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.000%
Target Compounds						
					Qvalue	
16) Acetone	2.374	43	111873	102.197	ug/l	98
18) Methyl Acetate	2.703	43	10092	4.142	ug/l	99
20) Methylene Chloride	2.794	84	22622	9.763	ug/l	93
25) 2-Butanone	4.562	43	11611	8.503	ug/l #	84
37) Ethyl Acetate	4.708	43	5542	1.591	ug/l #	90
43) Isopropyl Acetate	6.330	43	13477	2.359	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092225\
Data File : VX047701.D
Acq On : 22 Sep 2025 14:07
Operator : JC/MD
Sample : Q3133-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VNJ-238

Quant Time: Sep 23 01:34:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

