

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092225\
 Data File : VX047707.D
 Acq On : 22 Sep 2025 16:11
 Operator : JC/MD
 Sample : Q3157-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 20-DEAD-1

Quant Time: Sep 23 01:37:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 09/23/2025
 Supervised By : Mahesh Dadoda 09/24/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	155439	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	330407	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	332863	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	169366	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	136766	55.277	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 110.560%			
35) Dibromofluoromethane	5.373	113	107924	48.705	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 97.400%			
50) Toluene-d8	8.635	98	373596	48.725	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 97.460%			
62) 4-Bromofluorobenzene	11.061	95	159844	54.931	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 109.860%			
Target Compounds						
16) Acetone	2.380	43	19689	18.302	ug/l	97
18) Methyl Acetate	2.703	43	13838	5.780	ug/l	94
20) Methylene Chloride	2.788	84	8618	3.785	ug/l	91
37) Ethyl Acetate	4.715	43	7462m	2.152	ug/l	
43) Isopropyl Acetate	6.324	43	11374	2.000	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092225\
Data File : VX047707.D
Acq On : 22 Sep 2025 16:11
Operator : JC/MD
Sample : Q3157-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
20-DEAD-1

Quant Time: Sep 23 01:37:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 09/23/2025
Supervised By : Mahesh Dadoda 09/24/2025

