

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092319\
 Data File : VX012607.D
 Acq On : 23 Sep 2019 16:20
 Operator : JC/SP
 Sample : K5000-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANA

Quant Time: Sep 24 02:19:32 2019
 Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 23 12:27:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	174254	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	300516	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	250991	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	94249	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	126839	49.99	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.98%	
35) Dibromofluoromethane	5.49	113	90523	50.58	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.16%	
50) Toluene-d8	8.71	98	355340	51.56	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.12%	
62) 4-Bromofluorobenzene	11.14	95	120257	44.37	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.74%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092319\
Data File : VX012607.D
Acq On : 23 Sep 2019 16:20
Operator : JC/SP
Sample : K5000-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANA

Quant Time: Sep 24 02:19:32 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 23 12:27:01 2019
Response via : Initial Calibration

