

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092319\
 Data File : VX012611.D
 Acq On : 23 Sep 2019 17:54
 Operator : JC/SP
 Sample : K4930-05RE
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY038PS036-190916RE

Quant Time: Sep 24 02:26:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 23 12:27:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	166450	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	288017	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	231656	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	79310	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	119126	49.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.30%	
35) Dibromofluoromethane	5.49	113	85489	49.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.68%	
50) Toluene-d8	8.71	98	335541	50.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.60%	
62) 4-Bromofluorobenzene	11.13	95	108185	41.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.30%	
Target Compounds						
16) Acetone	2.44	43	2776	2.162	ug/l	99
27) cis-1,2-Dichloroethene	4.59	96	45660	19.848	ug/l	85
44) Trichloroethene	7.22	130	1440	0.690	ug/l	71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092319\
Data File : VX012611.D
Acq On : 23 Sep 2019 17:54
Operator : JC/SP
Sample : K4930-05RE
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY038PS036-190916RE

Quant Time: Sep 24 02:26:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 23 12:27:01 2019
Response via : Initial Calibration

