

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092319\
 Data File : VX012612.D
 Acq On : 23 Sep 2019 18:17
 Operator : JC/SP
 Sample : K4978-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z2-0164

Quant Time: Sep 24 02:30:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 23 12:27:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	171026	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	291274	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	240074	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	85996	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	123687	49.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.32%	
35) Dibromofluoromethane	5.49	113	88792	51.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.38%	
50) Toluene-d8	8.71	98	344979	51.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.28%	
62) 4-Bromofluorobenzene	11.13	95	114110	43.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.88%	
Target Compounds						
67) Ethyl Benzene	10.25	91	35206	3.840	ug/l	99
73) Isopropylbenzene	11.01	105	29162	4.164	ug/l	98
78) n-propylbenzene	11.35	91	31572	4.004	ug/l	98
85) sec-Butylbenzene	11.94	105	5466	0.829	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092319\
Data File : VX012612.D
Acq On : 23 Sep 2019 18:17
Operator : JC/SP
Sample : K4978-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z2-0164

Quant Time: Sep 24 02:30:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 23 12:27:01 2019
Response via : Initial Calibration

