

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092319\
 Data File : VX012616.D
 Acq On : 23 Sep 2019 19:51
 Operator : JC/SP
 Sample : K4978-02DL 50X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z2-0161DL

Quant Time: Sep 24 12:39:22 2019
 Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 23 12:27:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	169086	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	291836	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	248670	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	93875	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	121721	49.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.88%	
35) Dibromofluoromethane	5.49	113	87105	50.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.24%	
50) Toluene-d8	8.71	98	344067	51.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.82%	
62) 4-Bromofluorobenzene	11.14	95	120254	45.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.38%	
Target Compounds						
40) Benzene	6.14	78	302192	36.512	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092319\
Data File : VX012616.D
Acq On : 23 Sep 2019 19:51
Operator : JC/SP
Sample : K4978-02DL 50X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z2-0161DL

Quant Time: Sep 24 12:39:22 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 23 12:27:01 2019
Response via : Initial Calibration

