

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092320\
 Data File : VX018548.D
 Acq On : 23 Sep 2020 20:37
 Operator : JC/SP
 Sample : L4083-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE104D3-20200917

Quant Time: Sep 24 01:15:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 17 12:17:36 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	461855	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	727135	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	714372	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	354330	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	322705	49.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.78%	
35) Dibromofluoromethane	5.46	113	249540	49.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.28%	
50) Toluene-d8	8.70	98	891369	48.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.60%	
62) 4-Bromofluorobenzene	11.12	95	337786	46.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.98%	
Target Compounds						
3) Chloromethane	1.31	50	3102	0.605	ug/l	98
16) Acetone	2.43	43	8809	3.525	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092320\
Data File : VX018548.D
Acq On : 23 Sep 2020 20:37
Operator : JC/SP
Sample : L4083-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE104D3-20200917

Quant Time: Sep 24 01:15:58 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091720W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 17 12:17:36 2020
Response via : Initial Calibration

