

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092320\
 Data File : VX018550.D
 Acq On : 23 Sep 2020 21:22
 Operator : JC/SP
 Sample : L4083-12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE109D2-20200917

Quant Time: Sep 24 03:26:28 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 17 12:17:36 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	456641	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	732597	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	705630	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	370524	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	323622	50.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.20%	
35) Dibromofluoromethane	5.47	113	254502	50.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.50%	
50) Toluene-d8	8.70	98	889796	48.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.72%	
62) 4-Bromofluorobenzene	11.12	95	342248	46.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.52%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.38	101	5262	1.199	ug/l	97
44) Trichloroethene	7.19	130	179240	30.384	ug/l	96
64) Tetrachloroethene	9.32	164	3151	0.567	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092320\
Data File : VX018550.D
Acq On : 23 Sep 2020 21:22
Operator : JC/SP
Sample : L4083-12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE109D2-20200917

Quant Time: Sep 24 03:26:28 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091720W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 17 12:17:36 2020
Response via : Initial Calibration

