

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092320\
 Data File : VX018552.D
 Acq On : 23 Sep 2020 22:07
 Operator : JC/SP
 Sample : L4083-16
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB01-20200917

Quant Time: Sep 24 03:27:49 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 17 12:17:36 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	458302	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	740558	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	705077	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	368792	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	324480	50.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.10%	
35) Dibromofluoromethane	5.46	113	256583	50.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.24%	
50) Toluene-d8	8.70	98	908679	48.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.70%	
62) 4-Bromofluorobenzene	11.12	95	349696	47.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.52%	
Target Compounds						
16) Acetone	2.44	43	17656	7.121	ug/l	96
20) Methylene Chloride	2.84	84	4357	0.796	ug/l #	74
25) 2-Butanone	4.65	43	6537	1.693	ug/l	100
52) Toluene	8.77	92	99120	7.647	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092320\
Data File : VX018552.D
Acq On : 23 Sep 2020 22:07
Operator : JC/SP
Sample : L4083-16
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB01-20200917

Quant Time: Sep 24 03:27:49 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091720W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 17 12:17:36 2020
Response via : Initial Calibration

