

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092321\
 Data File : VX024465.D
 Acq On : 23 Sep 2021 21:35
 Operator : JC/MD
 Sample : M3833-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP01-20210920

Quant Time: Sep 24 07:39:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	141987	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	213881	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	202609	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	105970	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	102086	51.014	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 102.020%		
35) Dibromofluoromethane	5.397	113	78523	55.264	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 110.520%		
50) Toluene-d8	8.653	98	286534	54.791	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 109.580%		
62) 4-Bromofluorobenzene	11.085	95	110112	53.552	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 107.100%		
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.325	101	3624	2.585	ug/l	95
27) cis-1,2-Dichloroethene	4.495	96	984	0.570	ug/l	91
30) Chloroform	5.105	83	1466	0.492	ug/l #	81
44) Trichloroethene	7.135	130	696409	431.710	ug/l	99
64) Tetrachloroethene	9.281	164	1242	0.886	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092321\
Data File : VX024465.D
Acq On : 23 Sep 2021 21:35
Operator : JC/MD
Sample : M3833-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP01-20210920

Quant Time: Sep 24 07:39:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

