

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092321\
 Data File : VX024467.D
 Acq On : 23 Sep 2021 22:23
 Operator : JC/MD
 Sample : M3833-18
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE126D3-20210920

Quant Time: Sep 24 07:40:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	131007	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	222818	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	198409	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85247	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	97094	52.586	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.180%
35) Dibromofluoromethane	5.397	113	74937	50.625	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.260%
50) Toluene-d8	8.653	98	284716	52.260	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	104.520%
62) 4-Bromofluorobenzene	11.085	95	98074	45.785	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.560%
Target Compounds						
						Qvalue
44) Trichloroethene	7.129	130	3627	2.158	ug/l	92
64) Tetrachloroethene	9.281	164	1883	1.371	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092321\
Data File : VX024467.D
Acq On : 23 Sep 2021 22:23
Operator : JC/MD
Sample : M3833-18
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE126D3-20210920

Quant Time: Sep 24 07:40:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

