

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092321\
 Data File : VX024469.D
 Acq On : 23 Sep 2021 23:10
 Operator : JC/MD
 Sample : M3833-04DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW201D-20210919DL

Quant Time: Sep 24 07:40:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	128173	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	218961	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	191582	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	84709	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	93356	51.679	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 103.360%	
35) Dibromofluoromethane	5.397	113	74384	51.137	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.280%	
50) Toluene-d8	8.653	98	243903	45.557	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 91.120%#	
62) 4-Bromofluorobenzene	11.085	95	91548	43.491	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 86.980%	
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.172	85	1794	1.391	ug/l	93
7) Trichlorofluoromethane	1.892	101	43366	19.324	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.325	101	759	0.600	ug/l	91
12) 1,1-Dichloroethene	2.319	96	8231	7.005	ug/l	90
30) Chloroform	5.099	83	1758	0.653	ug/l #	80
32) 1,1,1-Trichloroethane	5.403	97	1831	0.766	ug/l #	46
44) Trichloroethene	7.135	130	24436	14.797	ug/l	95
64) Tetrachloroethene	9.275	164	19652	14.821	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092321\
Data File : VX024469.D
Acq On : 23 Sep 2021 23:10
Operator : JC/MD
Sample : M3833-04DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW201D-20210919DL

Quant Time: Sep 24 07:40:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

