

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092321\
 Data File : VX024470.D
 Acq On : 23 Sep 2021 23:33
 Operator : JC/MD
 Sample : M3833-09DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE103D2-20210920DL

Quant Time: Sep 24 07:41:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	132558	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	228160	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	209722	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	91424	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	99556	53.289	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	106.580%	
35) Dibromofluoromethane	5.391	113	75007	49.486	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	98.980%	
50) Toluene-d8	8.653	98	281338	50.431	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	100.860%	
62) 4-Bromofluorobenzene	11.085	95	93731	42.733	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	85.460%	
Target Compounds						Qvalue
44) Trichloroethene	7.135	130	79687	46.307	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092321\
Data File : VX024470.D
Acq On : 23 Sep 2021 23:33
Operator : JC/MD
Sample : M3833-09DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE103D2-20210920DL

Quant Time: Sep 24 07:41:08 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

