

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092321\
 Data File : VX024475.D
 Acq On : 24 Sep 2021 02:17
 Operator : JC/MD
 Sample : VX0923WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0923WBL02

Quant Time: Sep 24 07:42:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	146995	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	258765	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	246429	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	108129	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	106790	51.547	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 103.100%	
35) Dibromofluoromethane	5.397	113	86618	50.387	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.780%	
50) Toluene-d8	8.653	98	315202	49.819	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.640%	
62) 4-Bromofluorobenzene	11.085	95	115496	46.428	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 92.860%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092321\
Data File : VX024475.D
Acq On : 24 Sep 2021 02:17
Operator : JC/MD
Sample : VX0923WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0923WBL02

Quant Time: Sep 24 07:42:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

