

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092321\
 Data File : VX024478.D
 Acq On : 24 Sep 2021 03:28
 Operator : JC/MD
 Sample : M3829-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB01-20210917

Quant Time: Sep 24 07:43:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	149775	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	263534	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	249736	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	110142	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	107138	50.755	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 101.500%			
35) Dibromofluoromethane	5.397	113	88110	50.328	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 100.660%			
50) Toluene-d8	8.653	98	323983	50.280	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 100.560%			
62) 4-Bromofluorobenzene	11.085	95	119622	47.216	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 94.440%			
Target Compounds						
11) Tert butyl alcohol	2.971	59	12732	14.273	ug/l	98
16) Acetone	2.385	43	19782	12.580	ug/l	95
17) Carbon Disulfide	2.513	76	1353	0.608	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092321\
Data File : VX024478.D
Acq On : 24 Sep 2021 03:28
Operator : JC/MD
Sample : M3829-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB01-20210917

Quant Time: Sep 24 07:43:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

