

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092321\
 Data File : VX024481.D
 Acq On : 24 Sep 2021 04:38
 Operator : JC/MD
 Sample : M3829-18
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW203D-20210918

Quant Time: Sep 24 07:44:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	134198	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	232491	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	242367	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	101355	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	100129	52.940	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 105.880%			
35) Dibromofluoromethane	5.397	113	77982	50.490	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 100.980%			
50) Toluene-d8	8.653	98	289411	50.912	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 101.820%			
62) 4-Bromofluorobenzene	11.085	95	114428	51.197	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 102.400%			
Target Compounds					Qvalue	
12) 1,1-Dichloroethene	2.325	96	2296	1.866	ug/l	90
24) 1,1-Dichloroethane	3.611	63	9827	3.852	ug/l	94
27) cis-1,2-Dichloroethene	4.489	96	6270	3.840	ug/l	92
30) Chloroform	5.093	83	2280	0.809	ug/l	93
32) 1,1,1-Trichloroethane	5.397	97	3177	1.270	ug/l #	48
44) Trichloroethene	7.135	130	36337	20.723	ug/l	97
64) Tetrachloroethene	9.281	164	1554	0.926	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092321\
Data File : VX024481.D
Acq On : 24 Sep 2021 04:38
Operator : JC/MD
Sample : M3829-18
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW203D-20210918

Quant Time: Sep 24 07:44:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

