

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092321\
 Data File : VX024483.D
 Acq On : 24 Sep 2021 05:25
 Operator : JC/MD
 Sample : M3829-20
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW203D2-20210918

Quant Time: Sep 24 07:44:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	124192	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	234361	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	208634	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	94363	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	101561	58.024	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 116.040%			
35) Dibromofluoromethane	5.397	113	82725	53.134	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 106.260%			
50) Toluene-d8	8.653	98	272003	47.468	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 94.940%			
62) 4-Bromofluorobenzene	11.079	95	104249	46.270	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 92.540%			
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.319	96	1061	0.932	ug/l #	70
24) 1,1-Dichloroethane	3.617	63	8867	3.756	ug/l	96
27) cis-1,2-Dichloroethene	4.495	96	1125	0.745	ug/l	96
30) Chloroform	5.105	83	915	0.351	ug/l	91
44) Trichloroethene	7.135	130	1281	0.725	ug/l	73
64) Tetrachloroethene	9.275	164	494	0.342	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092321\
Data File : VX024483.D
Acq On : 24 Sep 2021 05:25
Operator : JC/MD
Sample : M3829-20
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW203D2-20210918

Quant Time: Sep 24 07:44:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

