

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092321\
 Data File : VX024485.D
 Acq On : 24 Sep 2021 06:12
 Operator : JC/MD
 Sample : M3829-24
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE123D2-20210918

Quant Time: Sep 24 07:45:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	144016	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	207785	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	206396	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	104901	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	104396	51.433	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 102.860%	
35) Dibromofluoromethane	5.397	113	84066	60.901	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 121.800%	
50) Toluene-d8	8.653	98	254304	50.055	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.120%	
62) 4-Bromofluorobenzene	11.085	95	107267	53.699	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 107.400%	
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.337	101	519	0.365	ug/l	91
44) Trichloroethene	7.129	130	4277	2.729	ug/l	88
64) Tetrachloroethene	9.275	164	2861	2.003	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092321\
Data File : VX024485.D
Acq On : 24 Sep 2021 06:12
Operator : JC/MD
Sample : M3829-24
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE123D2-20210918

Quant Time: Sep 24 07:45:10 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

