

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092321\
 Data File : VX024486.D
 Acq On : 24 Sep 2021 06:36
 Operator : JC/MD
 Sample : M3829-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE114D2-20210916

Quant Time: Sep 24 07:45:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	131785	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	207171	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	219080	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	99002	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	96987	52.218	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 104.440%	
35) Dibromofluoromethane	5.391	113	77966	56.649	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 113.300%	
50) Toluene-d8	8.653	98	264190	52.155	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 104.320%	
62) 4-Bromofluorobenzene	11.085	95	102397	51.413	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 102.820%	
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.172	85	1142	0.861	ug/l	93
9) 1,1,2-Trichlorotrifluo...	2.331	101	18565	14.268	ug/l	98
12) 1,1-Dichloroethene	2.325	96	1245	1.031	ug/l #	76
27) cis-1,2-Dichloroethene	4.501	96	2758	1.720	ug/l	90
30) Chloroform	5.105	83	1422	0.514	ug/l #	76
44) Trichloroethene	7.135	130	138888	88.886	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092321\
Data File : VX024486.D
Acq On : 24 Sep 2021 06:36
Operator : JC/MD
Sample : M3829-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE114D2-20210916

Quant Time: Sep 24 07:45:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

