

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092321\
 Data File : VX024493.D
 Acq On : 24 Sep 2021 09:20
 Operator : JC/MD
 Sample : M3829-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE120D3-20210917

Quant Time: Sep 24 16:35:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	122127	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	205067	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	198463	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	84104	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	91978	53.437	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 106.880%			
35) Dibromofluoromethane	5.397	113	70577	51.807	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 103.620%			
50) Toluene-d8	8.653	98	254026	50.663	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 101.320%			
62) 4-Bromofluorobenzene	11.085	95	92438	46.889	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 93.780%			
Target Compounds						Qvalue
44) Trichloroethene	7.129	130	43735	28.277	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092321\
Data File : VX024493.D
Acq On : 24 Sep 2021 09:20
Operator : JC/MD
Sample : M3829-14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE120D3-20210917

Quant Time: Sep 24 16:35:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

