

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092324\
 Data File : VX043106.D
 Acq On : 23 Sep 2024 11:05
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Sep 24 02:52:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092324W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 24 02:22:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	84705	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	162786	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	153297	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	76876	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	72092	51.238	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.480%
35) Dibromofluoromethane	5.385	113	56423	49.134	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.260%
50) Toluene-d8	8.647	98	201282	51.809	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.620%
62) 4-Bromofluorobenzene	11.079	95	81721	52.232	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	104.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092324\
Data File : VX043106.D
Acq On : 23 Sep 2024 11:05
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Sep 24 02:52:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092324W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 24 02:22:33 2024
Response via : Initial Calibration

