

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092324\
 Data File : VX043107.D
 Acq On : 23 Sep 2024 13:54
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX092324

Quant Time: Sep 24 02:53:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092324W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 24 02:22:33 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	86	0.00
2 T	Dichlorodifluoromethane	0.533	0.511	4.1	84	0.00
3 P	Chloromethane	0.544	0.508	6.6	86	0.00
4 C	Vinyl Chloride	0.553	0.520	6.0#	84	0.00
5 T	Bromomethane	0.318	0.311	2.2	73	0.00
6 T	Chloroethane	0.354	0.340	4.0	87	0.00
7 T	Trichlorofluoromethane	0.982	0.914	6.9	83	0.00
8 T	Diethyl Ether	0.330	0.348	-5.5	94	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.554	0.541	2.3	88	0.00
10 T	Methyl Iodide	0.658	0.655	0.5	86	0.00
11 T	Tert butyl alcohol	0.185	0.199	-7.6	96	0.00
12 CM	1,1-Dichloroethene	0.530	0.501	5.5#	85	0.00
13 T	Acrolein	0.024	0.024	0.0	132	0.00
14 T	Allyl chloride	0.984	0.947	3.8	89	0.00
15 T	Acrylonitrile	0.322	0.339	-5.3	95	0.00
16 T	Acetone	0.388	0.373	3.9	82	0.00
17 T	Carbon Disulfide	1.317	1.227	6.8	86	0.00
18 T	Methyl Acetate	1.045	1.144	-9.5	92	0.00
19 T	Methyl tert-butyl Ether	2.019	2.118	-4.9	94	0.00
20 T	Methylene Chloride	0.618	0.611	1.1	93	0.00
21 T	trans-1,2-Dichloroethene	0.548	0.533	2.7	90	0.00
22 T	Diisopropyl ether	1.842	1.880	-2.1	93	0.00
23 T	Vinyl Acetate	2.167	2.285	-5.4	96	0.00
24 P	1,1-Dichloroethane	1.065	1.061	0.4	90	0.00
25 T	2-Butanone	0.501	0.517	-3.2	91	0.00
26 T	2,2-Dichloropropane	1.078	1.050	2.6	89	0.00
27 T	cis-1,2-Dichloroethene	0.683	0.688	-0.7	92	0.00
28 T	Bromochloromethane	0.469	0.508	-8.3	101	0.00
29 T	Tetrahydrofuran	0.301	0.305	-1.3	93	0.00
30 C	Chloroform	1.175	1.176	-0.1#	91	0.00
31 T	Cyclohexane	0.817	0.784	4.0	86	0.00
32 T	1,1,1-Trichloroethane	1.113	1.064	4.4	87	0.00
33 S	1,2-Dichloroethane-d4	0.831	0.854	-2.8	92	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	88	0.00
35 S	Dibromofluoromethane	0.353	0.354	-0.3	90	0.00
36 T	1,1-Dichloropropene	0.433	0.409	5.5	88	0.00
37 T	Ethyl Acetate	0.544	0.539	0.9	96	0.00
38 T	Carbon Tetrachloride	0.558	0.513	8.1	86	0.00
39 T	Methylcyclohexane	0.544	0.505	7.2	87	0.00
40 TM	Benzene	1.279	1.258	1.6	91	0.00
41 T	Methacrylonitrile	0.271	0.303	-11.8	97	0.00
42 TM	1,2-Dichloroethane	0.543	0.553	-1.8	94	0.00
43 T	Isopropyl Acetate	0.909	0.938	-3.2	95	0.00
44 TM	Trichloroethene	0.350	0.335	4.3	88	0.00
45 C	1,2-Dichloropropane	0.319	0.326	-2.2#	93	0.00
46 T	Dibromomethane	0.268	0.272	-1.5	95	0.00
47 T	Bromodichloromethane	0.558	0.573	-2.7	93	0.00
48 T	Methyl methacrylate	0.447	0.458	-2.5	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092324\
 Data File : VX043107.D
 Acq On : 23 Sep 2024 13:54
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX092324

Quant Time: Sep 24 02:53:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092324W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 24 02:22:33 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.009	-12.5	96	0.00
50 S	Toluene-d8	1.193	1.178	1.3	89	0.00
51 T	4-Methyl-2-Pentanone	0.559	0.579	-3.6	95	0.00
52 CM	Toluene	0.824	0.827	-0.4#	91	0.00
53 T	t-1,3-Dichloropropene	0.573	0.591	-3.1	94	0.00
54 T	cis-1,3-Dichloropropene	0.572	0.597	-4.4	95	0.00
55 T	1,1,2-Trichloroethane	0.339	0.348	-2.7	94	0.00
56 T	Ethyl methacrylate	0.590	0.617	-4.6	95	0.00
57 T	1,3-Dichloropropane	0.577	0.585	-1.4	94	0.00
58 T	2-Chloroethyl Vinyl ether	0.239	0.247	-3.3	90	0.00
59 T	2-Hexanone	0.437	0.458	-4.8	93	0.00
60 T	Dibromochloromethane	0.443	0.458	-3.4	95	0.00
61 T	1,2-Dibromoethane	0.371	0.387	-4.3	96	0.00
62 S	4-Bromofluorobenzene	0.481	0.496	-3.1	92	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	88	0.00
64 T	Tetrachloroethene	0.362	0.329	9.1	84	0.00
65 PM	Chlorobenzene	1.079	1.063	1.5	92	0.00
66 T	1,1,1,2-Tetrachloroethane	0.398	0.394	1.0	91	0.00
67 C	Ethyl Benzene	1.866	1.812	2.9#	89	0.00
68 T	m/p-Xylenes	0.694	0.682	1.7	90	0.00
69 T	o-Xylene	0.696	0.694	0.3	93	0.00
70 T	Styrene	1.162	1.200	-3.3	93	0.00
71 P	Bromoform	0.350	0.368	-5.1	95	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	0.00
73 T	Isopropylbenzene	3.750	3.496	6.8	90	0.00
74 T	N-amyl acetate	1.762	1.784	-1.2	95	0.00
75 P	1,1,2,2-Tetrachloroethane	1.207	1.219	-1.0	97	0.00
76 T	1,2,3-Trichloropropane	1.063	1.091	-2.6	97	0.00
77 T	Bromobenzene	0.900	0.898	0.2	95	0.00
78 T	n-propylbenzene	4.254	4.056	4.7	90	0.00
79 T	2-Chlorotoluene	2.668	2.502	6.2	91	0.00
80 T	1,3,5-Trimethylbenzene	3.102	2.972	4.2	91	0.00
81 T	trans-1,4-Dichloro-2-butene	0.446	0.449	-0.7	96	0.00
82 T	4-Chlorotoluene	3.050	2.983	2.2	93	0.00
83 T	tert-Butylbenzene	3.197	3.048	4.7	93	0.00
84 T	1,2,4-Trimethylbenzene	3.087	2.999	2.9	91	0.00
85 T	sec-Butylbenzene	3.929	3.693	6.0	90	0.00
86 T	p-Isopropyltoluene	3.262	3.148	3.5	91	0.00
87 T	1,3-Dichlorobenzene	1.641	1.649	-0.5	95	0.00
88 T	1,4-Dichlorobenzene	1.646	1.625	1.3	95	0.00
89 T	n-Butylbenzene	2.843	2.867	-0.8	93	0.00
90 T	Hexachloroethane	0.650	0.629	3.2	90	0.00
91 T	1,2-Dichlorobenzene	1.616	1.602	0.9	95	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.326	0.325	0.3	95	0.00
93 T	1,2,4-Trichlorobenzene	1.002	1.043	-4.1	98	0.00
94 T	Hexachlorobutadiene	0.421	0.414	1.7	94	0.00
95 T	Naphthalene	3.483	3.576	-2.7	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092324\
Data File : VX043107.D
Acq On : 23 Sep 2024 13:54
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX092324

Quant Time: Sep 24 02:53:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092324W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 24 02:22:33 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.996	1.035	-3.9	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6