

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092325\
 Data File : VX047727.D
 Acq On : 23 Sep 2025 13:04
 Operator : JC/MD
 Sample : Q3141-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-184-EB-20250918

Quant Time: Sep 24 01:31:25 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	130435	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	270519	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	267539	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	132169	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	102139	49.195	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.400%
35) Dibromofluoromethane	5.373	113	85268	46.999	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.000%
50) Toluene-d8	8.635	98	300718	47.903	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.800%
62) 4-Bromofluorobenzene	11.061	95	126191	52.966	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.940%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.941	59	5744	24.841	ug/l	99
16) Acetone	2.374	43	69407	76.885	ug/l	98
25) 2-Butanone	4.556	43	11700	10.390	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092325\
Data File : VX047727.D
Acq On : 23 Sep 2025 13:04
Operator : JC/MD
Sample : Q3141-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-184-EB-20250918

Quant Time: Sep 24 01:31:25 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

