

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092325\
 Data File : VX047731.D
 Acq On : 23 Sep 2025 14:28
 Operator : JC/MD
 Sample : Q3141-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-184-GW-100-102

Quant Time: Sep 24 01:33:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	134925	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	279701	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	291553	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	150320	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	115961	53.994	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 107.980%		
35) Dibromofluoromethane	5.373	113	91274	48.658	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 97.320%		
50) Toluene-d8	8.635	98	316916	48.826	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 97.660%		
62) 4-Bromofluorobenzene	11.061	95	141454	57.423	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 114.840%		
Target Compounds						
						Qvalue
3) Chloromethane	1.301	50	618	0.337	ug/l	91
16) Acetone	2.374	43	11865	12.706	ug/l	98
17) Carbon Disulfide	2.520	76	1155	0.263	ug/l	100
30) Chloroform	5.087	83	2796	0.812	ug/l	97
47) Bromodichloromethane	7.812	83	1138	0.355	ug/l #	71
52) Toluene	8.702	92	1962	0.383	ug/l	96
60) Dibromochloromethane	9.506	129	1227	0.538	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092325\
Data File : VX047731.D
Acq On : 23 Sep 2025 14:28
Operator : JC/MD
Sample : Q3141-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-184-GW-100-102

Quant Time: Sep 24 01:33:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

