

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092325\
 Data File : VX047740.D
 Acq On : 23 Sep 2025 17:36
 Operator : JC/MD
 Sample : Q3097-22
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE134D1-20250912

Quant Time: Sep 24 01:37:11 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	132076	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	279533	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	278696	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	139922	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	107026	50.909	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 101.820%		
35) Dibromofluoromethane	5.379	113	90188	48.108	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 96.220%		
50) Toluene-d8	8.634	98	312762	48.215	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 96.420%		
62) 4-Bromofluorobenzene	11.061	95	131138	53.268	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 106.540%		
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.337	101	462	0.302	ug/l #	74
16) Acetone	2.380	43	1247	1.364	ug/l	97
27) cis-1,2-Dichloroethene	4.489	96	943	0.456	ug/l	71
44) Trichloroethene	7.116	130	10630	5.285	ug/l	98
64) Tetrachloroethene	9.256	164	990	0.545	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092325\
Data File : VX047740.D
Acq On : 23 Sep 2025 17:36
Operator : JC/MD
Sample : Q3097-22
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE134D1-20250912

Quant Time: Sep 24 01:37:11 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

