

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092325\
 Data File : VX047743.D
 Acq On : 23 Sep 2025 18:39
 Operator : JC/MD
 Sample : Q3151-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1402

Quant Time: Sep 24 01:38:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	131527	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	269092	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	269877	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	135495	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	102049	48.744	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.480%
35) Dibromofluoromethane	5.379	113	85203	47.212	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.420%
50) Toluene-d8	8.634	98	299724	47.998	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.000%
62) 4-Bromofluorobenzene	11.061	95	126188	53.246	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.500%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.337	101	939	0.617	ug/l	91
16) Acetone	2.379	43	3952	4.341	ug/l	95
17) Carbon Disulfide	2.520	76	1092	0.255	ug/l	97
29) Tetrahydrofuran	5.013	42	1997	2.924	ug/l #	83
44) Trichloroethene	7.116	130	1168	0.603	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092325\
Data File : VX047743.D
Acq On : 23 Sep 2025 18:39
Operator : JC/MD
Sample : Q3151-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1402

Quant Time: Sep 24 01:38:33 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

