

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092325\
 Data File : VX047750.D
 Acq On : 24 Sep 2025 04:52
 Operator : JC/MD
 Sample : Q3127-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE131D2-20250915

Quant Time: Sep 24 05:35:04 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.531	168	128787	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	6.739	114	267775	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	258758	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	127680	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.934	65	97517	47.570	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.140%
35) Dibromofluoromethane	5.367	113	83111	46.280	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.560%
50) Toluene-d8	8.634	98	288338	46.401	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.800%
62) 4-Bromofluorobenzene	11.061	95	119434	50.644	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.280%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	258012	173.209	ug/l	99
12) 1,1-Dichloroethene	2.325	96	3121	2.040	ug/l	94
27) cis-1,2-Dichloroethene	4.483	96	11038	5.471	ug/l	88
44) Trichloroethene	7.104	130	431582	223.976	ug/l	98
64) Tetrachloroethene	9.256	164	36170	21.443	ug/l	97
94) Hexachlorobutadiene	13.707	225	1100	1.178	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092325\
Data File : VX047750.D
Acq On : 24 Sep 2025 04:52
Operator : JC/MD
Sample : Q3127-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE131D2-20250915

Quant Time: Sep 24 05:35:04 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

