

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092325\
 Data File : VX047753.D
 Acq On : 24 Sep 2025 05:55
 Operator : JC/MD
 Sample : Q3127-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE132D6-20250916

Quant Time: Sep 24 07:10:22 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.537	168	147647	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	305020	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	291964	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	154008	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	120242	51.163	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 102.320%	
35) Dibromofluoromethane	5.373	113	99092	48.441	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 96.880%	
50) Toluene-d8	8.634	98	317405	44.842	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 89.680%#	
62) 4-Bromofluorobenzene	11.067	95	139890	52.075	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 104.140%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.337	101	65810	38.536	ug/l	99
12) 1,1-Dichloroethene	2.324	96	19473	11.101	ug/l	97
24) 1,1-Dichloroethane	3.605	63	7865	2.109	ug/l #	96
27) cis-1,2-Dichloroethene	4.489	96	10394	4.494	ug/l	90
30) Chloroform	5.086	83	6223	1.652	ug/l	99
38) Carbon Tetrachloride	5.671	117	6660	2.051	ug/l	97
44) Trichloroethene	7.116	130	3524733	1605.857	ug/l	94
55) 1,1,2-Trichloroethane	9.140	97	3082	1.429	ug/l	96
64) Tetrachloroethene	9.256	164	6273	3.296	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092325\
Data File : VX047753.D
Acq On : 24 Sep 2025 05:55
Operator : JC/MD
Sample : Q3127-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE132D6-20250916

Quant Time: Sep 24 07:10:22 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

