

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092325\
 Data File : VX047754.D
 Acq On : 24 Sep 2025 06:16
 Operator : JC/MD
 Sample : Q3127-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT101D2-20250915

Quant Time: Sep 24 07:10:55 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	139114	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	287200	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	275511	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	144556	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	109177	49.304	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.600%
35) Dibromofluoromethane	5.373	113	91968	47.748	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.500%
50) Toluene-d8	8.634	98	303571	45.549	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.100%#
62) 4-Bromofluorobenzene	11.061	95	132220	52.273	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.540%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.337	101	50996	31.693	ug/l	99
12) 1,1-Dichloroethene	2.325	96	16108	9.746	ug/l	94
24) 1,1-Dichloroethane	3.611	63	6635	1.889	ug/l	96
27) cis-1,2-Dichloroethene	4.483	96	7692	3.530	ug/l	93
30) Chloroform	5.092	83	4105	1.156	ug/l #	76
38) Carbon Tetrachloride	5.672	117	3459	1.131	ug/l	91
44) Trichloroethene	7.116	130	2850275	1379.149	ug/l	95
55) 1,1,2-Trichloroethane	9.140	97	2077	1.023	ug/l #	73
64) Tetrachloroethene	9.262	164	7900	4.399	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092325\
Data File : VX047754.D
Acq On : 24 Sep 2025 06:16
Operator : JC/MD
Sample : Q3127-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT101D2-20250915

Quant Time: Sep 24 07:10:55 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

