

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092325\
 Data File : VX047761.D
 Acq On : 24 Sep 2025 08:43
 Operator : JC/MD
 Sample : Q3141-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-184-EB-20250918

Quant Time: Sep 24 09:04:28 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	157285	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	293648	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	269079	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	126784	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	132556	52.947	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 105.900%			
35) Dibromofluoromethane	5.379	113	100689	51.128	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 102.260%			
50) Toluene-d8	8.635	98	337096	49.468	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 98.940%			
62) 4-Bromofluorobenzene	11.061	95	131474	50.837	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 101.680%			
Target Compounds						
11) Tert butyl alcohol	2.947	59	5373	19.270	ug/l	99
16) Acetone	2.374	43	70220	64.507	ug/l	94
25) 2-Butanone	4.556	43	10584	7.795	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092325\
Data File : VX047761.D
Acq On : 24 Sep 2025 08:43
Operator : JC/MD
Sample : Q3141-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-184-EB-20250918

Quant Time: Sep 24 09:04:28 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

