

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX092325\  
 Data File : VX047762.D  
 Acq On : 24 Sep 2025 09:04  
 Operator : JC/MD  
 Sample : Q3127-04  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 51 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 DUP07-20250915

Quant Time: Sep 25 01:19:39 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X091625W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 17 06:39:58 2025  
 Response via : Initial Calibration

| Compound                     | R.T. QIon |                | Response | Conc       | Units  | Dev(Min) |
|------------------------------|-----------|----------------|----------|------------|--------|----------|
| -----                        |           |                |          |            |        |          |
| Internal Standards           |           |                |          |            |        |          |
| 1) Pentafluorobenzene        | 5.544     | 168            | 121816   | 50.000     | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene      | 6.751     | 114            | 256774   | 50.000     | ug/l   | 0.00     |
| 63) Chlorobenzene-d5         | 10.037    | 117            | 253945   | 50.000     | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 12.006    | 152            | 124709   | 50.000     | ug/l   | 0.00     |
| System Monitoring Compounds  |           |                |          |            |        |          |
| 33) 1,2-Dichloroethane-d4    | 5.940     | 65             | 98601    | 50.851     | ug/l   | 0.00     |
| Spiked Amount                | 50.000    | Range 78 - 117 | Recovery | = 101.700% |        |          |
| 35) Dibromofluoromethane     | 5.379     | 113            | 82247    | 47.761     | ug/l   | 0.00     |
| Spiked Amount                | 50.000    | Range 75 - 124 | Recovery | = 95.520%  |        |          |
| 50) Toluene-d8               | 8.634     | 98             | 287969   | 48.328     | ug/l   | 0.00     |
| Spiked Amount                | 50.000    | Range 92 - 112 | Recovery | = 96.660%  |        |          |
| 62) 4-Bromofluorobenzene     | 11.067    | 95             | 118756   | 52.514     | ug/l   | 0.00     |
| Spiked Amount                | 50.000    | Range 83 - 123 | Recovery | = 105.020% |        |          |
| Target Compounds             |           |                |          |            |        |          |
|                              |           |                |          |            |        | Qvalue   |
| 9) 1,1,2-Trichlorotrifluo... | 2.337     | 101            | 3706     | 2.630      | ug/l   | 94       |
| 12) 1,1-Dichloroethene       | 2.325     | 96             | 965      | 0.667      | ug/l   | 97       |
| 19) Methyl tert-butyl Ether  | 3.111     | 73             | 5260     | 0.996      | ug/l   | 96       |
| 20) Methylene Chloride       | 2.788     | 84             | 504      | 0.282      | ug/l # | 83       |
| 24) 1,1-Dichloroethane       | 3.605     | 63             | 3227     | 1.049      | ug/l # | 95       |
| 27) cis-1,2-Dichloroethene   | 4.495     | 96             | 6922     | 3.627      | ug/l   | 81       |
| 30) Chloroform               | 5.080     | 83             | 4836     | 1.556      | ug/l # | 73       |
| 44) Trichloroethene          | 7.110     | 130            | 377625   | 204.371    | ug/l   | 95       |
| 64) Tetrachloroethene        | 9.256     | 164            | 22817    | 13.783     | ug/l   | 98       |
| -----                        |           |                |          |            |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX092325\  
Data File : VX047762.D  
Acq On : 24 Sep 2025 09:04  
Operator : JC/MD  
Sample : Q3127-04  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 51 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
DUP07-20250915

Quant Time: Sep 25 01:19:39 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X091625W.M  
Quant Title : SW846 8260  
QLast Update : Wed Sep 17 06:39:58 2025  
Response via : Initial Calibration

