

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092325\
 Data File : VX047729.D
 Acq On : 23 Sep 2025 13:46
 Operator : JC/MD
 Sample : Q3097-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB03-20250911

Quant Time: Sep 24 01:32:21 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.537	168	143699	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	290210	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	290111	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	146254	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	108356	47.372	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.740%
35) Dibromofluoromethane	5.379	113	91631	47.079	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.160%
50) Toluene-d8	8.635	98	319601	47.456	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.920%
62) 4-Bromofluorobenzene	11.067	95	136012	53.215	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.420%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092325\
Data File : VX047729.D
Acq On : 23 Sep 2025 13:46
Operator : JC/MD
Sample : Q3097-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB03-20250911


Quant Time: Sep 24 01:32:21 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.537 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX047729.D
Acq: 23 Sep 2025 13:46

Instrument :
MSVOA_X
ClientSampleId :
TB03-20250911







Tgt Ion:168 Resp: 143699
Ion Ratio Lower Upper
168 100
99 64.0 48.8 73.2



#33
1,2-Dichloroethane-d4
Concen: 47.372 ug/l
RT: 5.946 min Scan# 797
Delta R.T. 0.006 min
Lab File: VX047729.D
Acq: 23 Sep 2025 13:46


Tgt Ion: 65 Resp: 108356
Ion Ratio Lower Upper
65 100
67 52.4 0.0 103.0








#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.751 min Scan# 91
Delta R.T. 0.006 min
Lab File: VX047729.D
Acq: 23 Sep 2025 13:46

Instrument :
MSVOA_X
ClientSampleId :
TB03-20250911







Tgt Ion:114 Resp: 290210
Ion Ratio Lower Upper
114 100
63 22.7 0.0 44.0
88 16.2 0.0 34.2



#35
Dibromofluoromethane
Concen: 47.079 ug/l
RT: 5.379 min Scan# 704
Delta R.T. 0.006 min
Lab File: VX047729.D
Acq: 23 Sep 2025 13:46

Tgt Ion:113 Resp: 91631
Ion Ratio Lower Upper
113 100
111 102.1 80.9 121.3
192 17.9 14.2 21.4









#50
Toluene-d8
Concen: 47.456 ug/l
RT: 8.635 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX047729.D
Acq: 23 Sep 2025 13:46




Instrument :
MSVOA_X
ClientSampleId :
TB03-20250911


Tgt Ion: 98 Resp: 319601
Ion Ratio Lower Upper
98 100
100 66.4 53.0 79.4



#62
4-Bromofluorobenzene
Concen: 53.215 ug/l
RT: 11.067 min Scan# 1637
Delta R.T. 0.006 min
Lab File: VX047729.D
Acq: 23 Sep 2025 13:46

Tgt Ion: 95 Resp: 136012
Ion Ratio Lower Upper
95 100
174 71.3 0.0 141.6
176 69.4 0.0 138.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.037 min Scan# 1468
Delta R.T. -0.000 min
Lab File: VX047729.D
Acq: 23 Sep 2025 13:46

Instrument :
MSVOA_X
ClientSampleId :
TB03-20250911

Tgt Ion:117 Resp: 290111
Ion Ratio Lower Upper
117 100
82 59.1 47.8 71.6
119 33.3 25.2 37.8



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VX047729.D
Acq: 23 Sep 2025 13:46

Tgt Ion:152 Resp: 146254
Ion Ratio Lower Upper
152 100
115 61.9 44.9 134.5
150 156.3 0.0 352.0

