

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092419\
 Data File : VX012625.D
 Acq On : 24 Sep 2019 12:13
 Operator : JC/SP
 Sample : K4930-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY038PS027-190916

Quant Time: Sep 25 05:20:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 23 12:27:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	172202	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	292309	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	238947	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	87273	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	122266	48.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.52%	
35) Dibromofluoromethane	5.49	113	88189	50.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.32%	
50) Toluene-d8	8.71	98	343017	51.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.34%	
62) 4-Bromofluorobenzene	11.14	95	113393	43.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.04%	
Target Compounds						
16) Acetone	2.44	43	4128	3.108	ug/l	97
27) cis-1,2-Dichloroethene	4.58	96	95565	40.153	ug/l	86
44) Trichloroethene	7.21	130	7322	3.455	ug/l	97
64) Tetrachloroethene	9.33	164	4844	2.908	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092419\
Data File : VX012625.D
Acq On : 24 Sep 2019 12:13
Operator : JC/SP
Sample : K4930-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY038PS027-190916

Quant Time: Sep 25 05:20:09 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 23 12:27:01 2019
Response via : Initial Calibration

