

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092419\
 Data File : VX012626.D
 Acq On : 24 Sep 2019 12:36
 Operator : JC/SP
 Sample : K5000-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Sep 25 05:24:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 23 12:27:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	163549	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	282249	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	230501	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	77408	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	117131	49.18	ug/l	0.00
Spiked Amount			Recovery	=	98.36%	
35) Dibromofluoromethane	5.49	113	84177	50.08	ug/l	0.00
Spiked Amount			Recovery	=	100.16%	
50) Toluene-d8	8.71	98	331057	51.14	ug/l	0.00
Spiked Amount			Recovery	=	102.28%	
62) 4-Bromofluorobenzene	11.14	95	104797	41.17	ug/l	0.00
Spiked Amount			Recovery	=	82.34%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092419\
Data File : VX012626.D
Acq On : 24 Sep 2019 12:36
Operator : JC/SP
Sample : K5000-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Sep 25 05:24:37 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 23 12:27:01 2019
Response via : Initial Calibration

