

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092419\
 Data File : VX012631.D
 Acq On : 24 Sep 2019 14:33
 Operator : JC/SP
 Sample : K5013-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S49-0523

Quant Time: Sep 25 05:37:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 23 12:27:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	159266	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	275693	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	233041	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	85545	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	115565	49.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.66%	
35) Dibromofluoromethane	5.49	113	84352	51.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.76%	
50) Toluene-d8	8.71	98	320651	50.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.42%	
62) 4-Bromofluorobenzene	11.14	95	110886	44.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.20%	
Target Compounds						
16) Acetone	2.43	43	4954	4.033	ug/l	99
21) trans-1,2-Dichloroethene	3.15	96	1543	0.813	ug/l #	77
24) 1,1-Dichloroethane	3.69	63	5674	1.590	ug/l	98
27) cis-1,2-Dichloroethene	4.59	96	2508	1.139	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092419\
Data File : VX012631.D
Acq On : 24 Sep 2019 14:33
Operator : JC/SP
Sample : K5013-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S49-0523

Quant Time: Sep 25 05:37:05 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 23 12:27:01 2019
Response via : Initial Calibration

