

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092419\
 Data File : VX012635.D
 Acq On : 24 Sep 2019 16:07
 Operator : JC/SP
 Sample : K5013-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S49-0544

Quant Time: Sep 25 06:07:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 23 12:27:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	154921	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	261581	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	211651	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	71758	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	109865	48.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.40%	
35) Dibromofluoromethane	5.49	113	78618	50.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.94%	
50) Toluene-d8	8.71	98	305178	50.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.74%	
62) 4-Bromofluorobenzene	11.14	95	97887	41.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.00%	
Target Compounds						
4) Vinyl Chloride	1.40	62	2265	1.139	ug/l #	78
21) trans-1,2-Dichloroethene	3.16	96	1960	1.062	ug/l	86
24) 1,1-Dichloroethane	3.69	63	3803	1.096	ug/l	96
27) cis-1,2-Dichloroethene	4.59	96	4498	2.101	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092419\
Data File : VX012635.D
Acq On : 24 Sep 2019 16:07
Operator : JC/SP
Sample : K5013-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S49-0544

Quant Time: Sep 25 06:07:19 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 23 12:27:01 2019
Response via : Initial Calibration

