

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX092419\
 Data File : VX012620.D
 Acq On : 24 Sep 2019 09:59
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 25 04:50:59 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 23 12:27:01 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	92	0.00
2 T	Dichlorodifluoromethane	0.562	0.561	0.2	98	0.00
3 P	Chloromethane	0.646	0.635	1.7	96	0.00
4 C	Vinyl Chloride	0.642	0.641	0.2#	96	0.00
5 T	Bromomethane	0.252	0.241	4.4	100	0.00
6 T	Chloroethane	0.433	0.404	6.7	95	0.00
7 T	Trichlorofluoromethane	0.901	0.924	-2.6	100	0.00
8 T	Diethyl Ether	0.369	0.358	3.0	94	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.549	0.578	-5.3	101	0.00
10 T	Methyl Iodide	0.517	0.573	-10.8	99	0.00
11 T	Tert butyl alcohol	0.182	0.172	5.5	88	0.00
12 CM	1,1-Dichloroethene	0.538	0.549	-2.0#	94	0.00
13 T	Acrolein	0.094	0.092	2.1	86	0.00
14 T	Allyl chloride	1.067	1.077	-0.9	95	0.00
15 T	Acrylonitrile	0.369	0.362	1.9	91	0.00
16 T	Acetone	0.386	0.453	-17.4	108	0.00
17 T	Carbon Disulfide	1.507	1.484	1.5	94	0.00
18 T	Methyl Acetate	0.908	0.865	4.7	92	0.00
19 T	Methyl tert-butyl Ether	2.035	2.035	0.0	94	0.00
20 T	Methylene Chloride	0.652	0.633	2.9	95	0.00
21 T	trans-1,2-Dichloroethene	0.596	0.594	0.3	95	0.00
22 T	Diisopropyl ether	2.061	2.060	0.0	95	0.00
23 T	Vinyl Acetate	1.807	1.828	-1.2	93	0.00
24 P	1,1-Dichloroethane	1.120	1.115	0.4	94	0.00
25 T	2-Butanone	0.553	0.588	-6.3	99	0.00
26 T	2,2-Dichloropropane	0.996	1.018	-2.2	97	0.00
27 T	cis-1,2-Dichloroethene	0.691	0.679	1.7	93	0.00
28 T	Bromochloromethane	0.407	0.420	-3.2	97	0.00
29 T	Tetrahydrofuran	0.337	0.330	2.1	90	0.00
30 C	Chloroform	1.146	1.130	1.4#	96	0.00
31 T	Cyclohexane	1.013	1.051	-3.8	99	0.00
32 T	1,1,1-Trichloroethane	1.016	1.033	-1.7	96	0.00
33 S	1,2-Dichloroethane-d4	0.728	0.735	-1.0	96	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	89	0.00
35 S	Dibromofluoromethane	0.298	0.317	-6.4	98	0.00
36 T	1,1-Dichloropropene	0.482	0.493	-2.3	98	0.00
37 T	Ethyl Acetate	0.563	0.560	0.5	90	0.00
38 T	Carbon Tetrachloride	0.490	0.505	-3.1	95	0.00
39 T	Methylcyclohexane	0.580	0.606	-4.5	99	0.00
40 TM	Benzene	1.418	1.430	-0.8	94	0.00
41 T	Methacrylonitrile	0.316	0.321	-1.6	94	0.00
42 TM	1,2-Dichloroethane	0.538	0.548	-1.9	95	0.00
43 T	Isopropyl Acetate	0.942	0.929	1.4	92	0.00
44 TM	Trichloroethene	0.363	0.368	-1.4	96	0.00
45 C	1,2-Dichloropropane	0.363	0.372	-2.5#	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX092419\
 Data File : VX012620.D
 Acq On : 24 Sep 2019 09:59
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 25 04:50:59 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 23 12:27:01 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.247	0.246	0.4	94	0.00
47 T	Bromodichloromethane	0.494	0.504	-2.0	93	0.00
48 T	Methyl methacrylate	0.452	0.457	-1.1	92	0.00
49 T	1,4-Dioxane	0.010	0.011	-10.0	93	0.00
50 S	Toluene-d8	1.147	1.175	-2.4	96	0.00
51 T	4-Methyl-2-Pentanone	0.582	0.578	0.7	91	0.00
52 CM	Toluene	0.891	0.919	-3.1#	96	0.00
53 T	t-1,3-Dichloropropene	0.547	0.574	-4.9	95	0.00
54 T	cis-1,3-Dichloropropene	0.597	0.618	-3.5	94	0.00
55 T	1,1,2-Trichloroethane	0.354	0.360	-1.7	94	0.00
56 T	Ethyl methacrylate	0.595	0.604	-1.5	92	0.00
57 T	1,3-Dichloropropane	0.620	0.623	-0.5	94	0.00
58 T	2-Chloroethyl Vinyl ether	0.267	0.293	-9.7	94	0.00
59 T	2-Hexanone	0.445	0.474	-6.5	95	0.00
60 T	Dibromochloromethane	0.359	0.377	-5.0	93	0.00
61 T	1,2-Dibromoethane	0.369	0.381	-3.3	94	0.00
62 S	4-Bromofluorobenzene	0.451	0.470	-4.2	98	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	87	0.00
64 T	Tetrachloroethene	0.349	0.381	-9.2	100	0.00
65 PM	Chlorobenzene	1.053	1.084	-2.9	94	0.00
66 T	1,1,1,2-Tetrachloroethane	0.381	0.406	-6.6	95	0.00
67 C	Ethyl Benzene	1.910	2.038	-6.7#	96	0.00
68 T	m/p-Xylenes	0.715	0.761	-6.4	95	0.00
69 T	o-Xylene	0.692	0.738	-6.6	96	0.00
70 T	Styrene	1.160	1.276	-10.0	96	0.00
71 P	Bromoform	0.266	0.291	-9.4	95	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	0.00
73 T	Isopropylbenzene	4.072	4.122	-1.2	96	0.00
74 T	N-amyl acetate	1.943	1.979	-1.9	93	0.00
75 P	1,1,2,2-Tetrachloroethane	1.349	1.349	0.0	94	0.00
76 T	1,2,3-Trichloropropane	1.323	1.340	-1.3	93	0.00
77 T	Bromobenzene	0.922	0.923	-0.1	95	0.00
78 T	n-propylbenzene	4.585	4.699	-2.5	97	0.00
79 T	2-Chlorotoluene	2.860	2.814	1.6	95	0.00
80 T	1,3,5-Trimethylbenzene	3.415	3.531	-3.4	97	0.00
81 T	trans-1,4-Dichloro-2-butene	0.462	0.479	-3.7	94	0.00
82 T	4-Chlorotoluene	3.260	3.383	-3.8	98	0.00
83 T	tert-Butylbenzene	3.294	3.311	-0.5	96	0.00
84 T	1,2,4-Trimethylbenzene	3.445	3.541	-2.8	96	0.00
85 T	sec-Butylbenzene	3.832	3.983	-3.9	97	0.00
86 T	p-Isopropyltoluene	3.457	3.711	-7.3	99	0.00
87 T	1,3-Dichlorobenzene	1.659	1.697	-2.3	96	0.00
88 T	1,4-Dichlorobenzene	1.676	1.725	-2.9	97	0.00
89 T	n-Butylbenzene	3.047	3.189	-4.7	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX092419\
Data File : VX012620.D
Acq On : 24 Sep 2019 09:59
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Sep 25 04:50:59 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 23 12:27:01 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.616	0.630	-2.3	95	0.00
91 T	1,2-Dichlorobenzene	1.661	1.699	-2.3	95	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.337	0.351	-4.2	91	0.00
93 T	1,2,4-Trichlorobenzene	1.013	1.068	-5.4	97	0.00
94 T	Hexachlorobutadiene	0.436	0.478	-9.6	103	0.00
95 T	Naphthalene	3.664	3.908	-6.7	93	0.00
96 T	1,2,3-Trichlorobenzene	1.014	1.072	-5.7	95	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6