

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092420\
 Data File : VX018560.D
 Acq On : 24 Sep 2020 10:33
 Operator : JC/SP
 Sample : VX0924WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0924WBL01

Quant Time: Sep 25 01:38:33 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 17 12:17:36 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	470601	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	764348	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	727529	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	348906	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	338117	50.79	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.58%	
35) Dibromofluoromethane	5.46	113	266437	50.42	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.84%	
50) Toluene-d8	8.70	98	914340	47.63	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.26%	
62) 4-Bromofluorobenzene	11.12	95	332942	43.60	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.20%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092420\
Data File : VX018560.D
Acq On : 24 Sep 2020 10:33
Operator : JC/SP
Sample : VX0924WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0924WBL01

Quant Time: Sep 25 01:38:33 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091720W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 17 12:17:36 2020
Response via : Initial Calibration

