

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092420\
 Data File : VX018562.D
 Acq On : 24 Sep 2020 11:25
 Operator : JC/SP
 Sample : L4083-16
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB01-20200917

Quant Time: Sep 25 01:42:53 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 17 12:17:36 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	479683	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	754781	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	728579	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	353761	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	332059	48.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.86%	
35) Dibromofluoromethane	5.47	113	258708	49.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.16%	
50) Toluene-d8	8.70	98	908854	47.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.88%	
62) 4-Bromofluorobenzene	11.12	95	351016	46.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.10%	
Target Compounds						
16) Acetone	2.43	43	19523	7.523	ug/l	Qvalue 100
52) Toluene	8.77	92	96295	7.289	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092420\
Data File : VX018562.D
Acq On : 24 Sep 2020 11:25
Operator : JC/SP
Sample : L4083-16
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB01-20200917

Quant Time: Sep 25 01:42:53 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091720W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 17 12:17:36 2020
Response via : Initial Calibration

