

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092420\
 Data File : VX018565.D
 Acq On : 24 Sep 2020 12:32
 Operator : JC/SP
 Sample : L4118-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE132D2-20200918

Quant Time: Sep 25 01:49:47 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 17 12:17:36 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	470003	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	740633	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	732455	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	364974	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	329125	49.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.00%	
35) Dibromofluoromethane	5.47	113	258622	50.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.02%	
50) Toluene-d8	8.70	98	911047	48.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.94%	
62) 4-Bromofluorobenzene	11.12	95	337667	45.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.26%	
Target Compounds						
2) Dichlorodifluoromethane	1.18	85	7595	1.704	ug/l	95
9) 1,1,2-Trichlorotrifluoroet	2.37	101	13374	2.961	ug/l	97
16) Acetone	2.43	43	7897	3.106	ug/l	95
44) Trichloroethene	7.19	130	10195	1.709	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092420\
Data File : VX018565.D
Acq On : 24 Sep 2020 12:32
Operator : JC/SP
Sample : L4118-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE132D2-20200918

Quant Time: Sep 25 01:49:47 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091720W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 17 12:17:36 2020
Response via : Initial Calibration

