

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092420\
 Data File : VX018566.D
 Acq On : 24 Sep 2020 12:54
 Operator : JC/SP
 Sample : L4118-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE132D3-20200918

Quant Time: Sep 25 01:52:07 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 17 12:17:36 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	448798	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	726328	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	706808	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	353058	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	319357	50.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.60%	
35) Dibromofluoromethane	5.47	113	255250	50.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.66%	
50) Toluene-d8	8.70	98	892896	48.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.88%	
62) 4-Bromofluorobenzene	11.12	95	339121	46.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.46%	
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.18	85	2896	0.681	ug/l #	68
9) 1,1,2-Trichlorotrifluoroet	2.37	101	17103	3.966	ug/l	95
12) 1,1-Dichloroethene	2.36	96	2771	0.638	ug/l #	89
27) cis-1,2-Dichloroethene	4.56	96	5649	0.998	ug/l	87
44) Trichloroethene	7.19	130	490036	83.786	ug/l	97
64) Tetrachloroethene	9.32	164	13089	2.352	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092420\
Data File : VX018566.D
Acq On : 24 Sep 2020 12:54
Operator : JC/SP
Sample : L4118-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE132D3-20200918

Quant Time: Sep 25 01:52:07 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091720W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 17 12:17:36 2020
Response via : Initial Calibration

