

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092420\
 Data File : VX018574.D
 Acq On : 24 Sep 2020 15:54
 Operator : JC/SP
 Sample : L4118-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW203D2-20200919

Quant Time: Sep 25 02:13:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 17 12:17:36 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	440702	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	720613	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	699646	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	347020	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	324730	52.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.18%	
35) Dibromofluoromethane	5.47	113	250505	50.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.56%	
50) Toluene-d8	8.70	98	884376	48.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.72%	
62) 4-Bromofluorobenzene	11.12	95	331768	46.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.16%	
Target Compounds						
12) 1,1-Dichloroethene	2.36	96	2648	0.620	ug/l	93
24) 1,1-Dichloroethane	3.67	63	17023	1.899	ug/l	97
27) cis-1,2-Dichloroethene	4.57	96	2986	0.537	ug/l	79
44) Trichloroethene	7.19	130	8097	1.395	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092420\
Data File : VX018574.D
Acq On : 24 Sep 2020 15:54
Operator : JC/SP
Sample : L4118-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW203D2-20200919

Quant Time: Sep 25 02:13:29 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091720W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 17 12:17:36 2020
Response via : Initial Calibration

